

ST VINCENT MEDICAL CENTER
LABORATORY(LAB# 7300)
2131 WEST THIRD STREET
LOS ANGELES CA USA 90057

Abnormal Fibrinogen (mg/dL)										
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Abnormal Fbg Control - Lot# LFC										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE THROMBIN									
Mean	105.2	109.0	-	-	-	-	109.0	106.0	106.0	101.0
SD	5.05	2.30	N	N	N	N	N	3.40	3.10	3.00
CV	4.8	2.1	-	-	-	-	N	3.2	2.9	3.0
# Points	123	6	-	-	-	-	4	21	21	21
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE THROMBIN									
Avg Mean	110.1	110.7	-	-	-	-	108.8	109.1	110.0	109.4
Peers	69	32	0	0	0	0	40	40	40	34
SDI	-1.2	-0.3	-	-	-	-	0.1	-0.8	-0.8	-2.6
CVI	0.9	0.5	-	-	-	-	1.2	0.7	0.6	0.7
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Abnormal Fbg Control - Lot# LFC										
Your Lab	SYSMEX CA 1500 - 2 , LIGHT SCATTER, DADE THROMBIN									
Mean	101.3	99.5	-	102.0	-	-	100.0	-	-	-
SD	4.34	4.50	N	3.00	N	N	2.20	N	N	N
CV	4.3	4.5	-	2.9	-	-	2.2	-	-	-
# Points	106	12	-	11	-	-	6	-	-	-
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE THROMBIN									
Avg Mean	110.1	110.7	-	111.1	-	-	108.8	-	-	-
Peers	69	32	0	35	0	0	40	0	0	0
SDI	-2.2	-2.2	-	-2.0	-	-	-2.2	-	-	-
CVI	0.8	1.0	-	0.7	-	-	0.5	-	-	-
APTT (seconds)										
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 1 - Lot# 548015										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE ACTIN FSL									
Mean	25.9	-	25.8	25.8	25.8	25.7	25.8	26.0	26.0	26.0
SD	0.42	N	0.29	0.34	0.36	0.30	0.35	0.34	0.41	0.49
CV	1.6	-	1.1	1.3	1.4	1.2	1.4	1.3	1.6	1.9
# Points	845	-	49	34	28	32	48	89	100	93
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE ACTIN FSL									
Avg Mean	25.5	25.3	25.4	25.5	25.6	25.6	25.6	25.5	25.6	25.5
Peers	103	38	52	65	75	80	77	77	75	74
SDI	0.8	-	1.0	0.6	0.5	0.2	0.5	1.0	0.9	1.0
CVI	0.9	-	0.7	0.8	0.8	0.7	0.8	0.8	0.9	1.2
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 3 - Lot# 548459										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE ACTIN FSL									
Mean	59.8	60.2	60.4	60.6	60.3	60.0	60.3	60.6	60.1	59.9
SD	1.35	0.90	0.97	0.98	0.66	1.05	1.50	1.48	1.47	1.44
CV	2.3	1.5	1.6	1.6	1.1	1.8	2.5	2.4	2.4	2.4
# Points	869	35	49	33	28	32	48	90	98	87
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE ACTIN FSL									
Avg Mean	59.7	59.9	60.0	60.1	60.1	60.2	59.9	59.8	59.7	59.6
Peers	102	40	51	64	74	79	76	76	74	74
SDI	0.1	0.4	0.6	0.5	0.1	-0.1	0.3	0.6	0.3	0.2
CVI	1.0	0.7	0.7	0.8	0.5	0.8	1.2	1.1	1.2	1.2
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 3 - Lot# 548459										
Your Lab	SYSMEX CA 1500 - 2 , LIGHT SCATTER, DADE ACTIN FSL									
Mean	60.3	61.0	61.2	-	60.5	61.2	60.8	60.0	59.5	59.4
SD	1.88	1.32	1.62	N	4.12	1.70	1.49	0.96	0.67	0.59
CV	3.1	2.2	2.6	-	6.8	2.8	2.5	1.6	1.1	1.0
# Points	918	94	129	-	75	51	70	31	31	30
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE ACTIN FSL									
Avg Mean	59.7	59.9	60.0	-	60.1	60.2	59.9	59.8	59.7	59.6
Peers	102	40	51	0	74	79	76	76	74	74
SDI	0.6	1.6	1.8	-	0.3	0.7	0.7	0.2	-0.2	-0.2
CVI	1.3	1.1	1.2	-	3.0	1.3	1.2	0.7	0.6	0.5

ST VINCENT MEDICAL CENTER
LABORATORY(LAB# 7300)
2131 WEST THIRD STREET
LOS ANGELES CA USA 90057

Claus Fibrinogen (mg/dL)										
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 1 - Lot# 548015										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE THROMBIN									
Mean	247.2	-	247.0	-	-	-	245.0	248.0	245.0	240.0
SD	8.10	N	N	N	N	N	N	9.40	7.10	7.00
CV	3.3	-	N	-	-	-	N	3.8	2.9	2.9
# Points	120	-	1	-	-	-	4	21	24	20
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE THROMBIN									
Avg Mean	257.7	256.5	258.6	-	-	-	256.4	258.2	259.4	258.9
Peers	74	31	37	0	0	0	46	44	42	39
SDI	-1.5	-	-1.1	-	-	-	-1.3	-1.1	-1.5	-2.3
CVI	0.8	-	0.0	-	-	-	0.3	1.0	0.7	0.8
Prothrombin Time (seconds)										
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 1 - Lot# 548015										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE INNOVIN									
Mean	10.7	-	10.6	10.7	10.7	10.7	10.5	10.7	10.7	10.8
SD	0.22	N	0.15	0.12	0.12	0.14	0.36	0.23	0.17	0.23
CV	2.0	-	1.4	1.1	1.1	1.3	3.4	2.1	1.6	2.1
# Points	844	-	49	34	34	17	55	92	100	88
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE INNOVIN									
Avg Mean	10.8	10.8	10.8	10.8	10.8	10.8	10.8	10.8	10.8	10.8
Peers	103	38	52	66	74	80	78	78	77	77
SDI	-0.7	-	-1.3	-0.5	-0.5	-0.5	-1.9	-0.8	-0.9	-0.3
CVI	1.1	-	0.7	0.7	0.7	0.7	1.6	1.2	0.9	1.2
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 3 - Lot# 548459										
Your Lab	SYSMEX CA 1500 - 1 , LIGHT SCATTER, DADE INNOVIN									
Mean	53.1	55.2	53.6	53.5	53.5	54.3	53.2	53.6	53.0	51.8
SD	2.24	1.82	1.74	1.57	1.57	1.30	1.96	2.44	2.16	2.51
CV	4.2	3.3	3.2	2.9	2.9	2.4	3.7	4.6	4.1	4.8
# Points	874	35	50	33	33	32	48	90	97	87
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE INNOVIN									
Avg Mean	53.2	53.0	52.9	52.9	53.0	53.1	53.3	53.4	53.5	53.6
Peers	104	40	51	66	75	80	78	78	77	76
SDI	-0.1	2.4	0.9	0.6	0.5	1.0	-0.1	0.2	-0.4	-1.5
CVI	1.3	1.2	1.3	1.1	1.0	0.8	1.2	1.4	1.4	1.6
	LTD	Oct	Sep	Aug	Jul	Jun	May	Apr	Mar	Feb
Ci-Trol 3 - Lot# 548459										
Your Lab	SYSMEX CA 1500 - 2 , LIGHT SCATTER, DADE INNOVIN									
Mean	54.0	55.0	53.7	53.8	53.1	54.6	54.6	54.5	54.3	53.9
SD	2.53	2.27	2.11	1.93	5.21	2.48	1.78	1.46	1.66	1.61
CV	4.7	4.1	3.9	3.6	9.8	4.5	3.3	2.7	3.1	3.0
# Points	1074	94	144	96	75	95	70	31	32	30
Test System	SYSMEX CA 1500/7000, LIGHT SCATTER, DADE INNOVIN									
Avg Mean	53.2	53.0	52.9	52.9	53.0	53.1	53.3	53.4	53.5	53.6
Peers	104	40	51	66	75	80	78	78	77	76
SDI	0.6	2.2	1.0	0.9	0.1	1.2	1.0	0.9	0.6	0.2
CVI	1.5	1.5	1.5	1.3	3.4	1.6	1.1	0.9	1.0	1.0

QAP Reports are designed primarily to summarize data on Siemens Hemostasis Controls. The statistics are calculations from user input and are provided as a service to customers. No claim is made as to the validity of values, tests and methods beyond those specifically identified on Siemens Hemostasis Control inserts and product literature.